

## Datalog

Current Event:19/09/09 09:19

### Summary

-----  
Unit Name           MiniRAE 3000(PGM-7320)  
Unit SN             592-911156  
Unit Firmware Ver   V2.16  
-----

Running Mode        Hygiene Mode  
Datalog Mode        Auto  
Diagnostic Mode     No  
Stop Reason         Power Down  
-----

Site ID             RAE00000  
User ID             00000001  
-----

Begin               9/9/2019 09:19:45  
End                  9/9/2019 15:05:02  
Sample Period(s)    900  
Number of Records   23  
-----

Sensor              PID(ppm)  
Sensor SN           S023030015Q9  
Measure Type        Min; Avg; Max  
Span                100.0  
Span 2              1000.0  
Low Alarm           50.0  
High Alarm          100.0  
Over Alarm          15000.0  
STEL Alarm          100.0  
TWA Alarm           50.0  
Measurement Gas     Isobutylene  
Calibration Time    8/29/2019 16:41  
Peak                N/A  
Min                 N/A  
Average             N/A

Sheet

Index	Date/Time	PID(ppm) (Min)	PID(ppm) (Avg)	PID(ppm) (Max)
001	9/9/2019 09:34:45	0.0	0.0	0.1
002	9/9/2019 09:49:45	0.0	0.1	0.1
003	9/9/2019 10:04:45	0.0	0.1	0.1
004	9/9/2019 10:19:45	0.0	0.0	0.1
005	9/9/2019 10:34:45	0.0	0.1	0.1
006	9/9/2019 10:49:45	0.1	0.1	0.1
007	9/9/2019 11:04:45	0.1	0.2	0.7
008	9/9/2019 11:19:45	0.5	0.9	1.0
009	9/9/2019 11:34:45	0.5	0.6	0.8
010	9/9/2019 11:49:45	0.4	0.5	0.6
011	9/9/2019 12:04:45	0.4	0.4	0.5
012	9/9/2019 12:19:45	0.1	0.3	0.4
013	9/9/2019 12:34:45	0.1	0.1	0.3
014	9/9/2019 12:49:45	0.2	0.2	0.3
015	9/9/2019 13:04:45	0.3	0.3	0.4
016	9/9/2019 13:19:45	0.3	0.4	0.4
017	9/9/2019 13:34:45	0.4	0.4	0.5
018	9/9/2019 13:49:45	0.2	0.3	0.4
019	9/9/2019 14:04:45	0.2	0.3	0.4
020	9/9/2019 14:19:45	0.3	0.4	0.4
021	9/9/2019 14:34:45	0.2	0.3	0.4
022	9/9/2019 14:49:45	0.1	0.2	0.3
023	9/9/2019 15:04:45	0.0	0.3	0.4
Peak		0.5	0.9	1.0
Min		0.0	0.0	0.1
Average		0.2	0.3	0.4