

Datalog

Current Event:19/09/06 08:53

Summary

Unit Name MiniRAE 3000(PGM-7320)
Unit SN 592-911156
Unit Firmware Ver V2.16

Running Mode Hygiene Mode
Datalog Mode Auto
Diagnostic Mode No
Stop Reason Power Down

Site ID RAE00000
User ID 00000001

Begin 9/6/2019 08:53:06
End 9/6/2019 14:15:32
Sample Period(s) 900
Number of Records 21

Sensor PID(ppm)
Sensor SN S023030015Q9
Measure Type Min; Avg; Max
Span 100.0
Span 2 1000.0
Low Alarm 50.0
High Alarm 100.0
Over Alarm 15000.0
STEL Alarm 100.0
TWA Alarm 50.0
Measurement Gas Isobutylene
Calibration Time 8/29/2019 16:41
Peak N/A
Min N/A
Average N/A

Sheet

Index	Date/Time	PID(ppm) (Min)	PID(ppm) (Avg)	PID(ppm) (Max)
001	9/6/2019 09:08:06	0.0	0.7	4.0
002	9/6/2019 09:23:06	0.0	1.3	13.3
003	9/6/2019 09:38:06	1.0	1.5	2.5
004	9/6/2019 09:53:06	1.0	1.3	1.5
005	9/6/2019 10:08:06	1.1	1.3	1.5
006	9/6/2019 10:23:06	0.8	1.0	1.2
007	9/6/2019 10:38:06	0.8	0.9	1.6
008	9/6/2019 10:53:06	0.9	1.2	1.9
009	9/6/2019 11:08:06	1.2	1.5	2.0
010	9/6/2019 11:23:06	1.1	1.7	2.9
011	9/6/2019 11:38:06	1.6	1.8	2.1
012	9/6/2019 11:53:06	1.8	2.3	3.0
013	9/6/2019 12:08:06	2.9	3.5	3.9
014	9/6/2019 12:23:06	3.2	4.5	5.7
015	9/6/2019 12:38:06	4.9	5.4	6.4
016	9/6/2019 12:53:06	4.9	5.3	5.6
017	9/6/2019 13:08:06	4.2	5.0	5.6
018	9/6/2019 13:23:06	3.7	3.9	4.4
019	9/6/2019 13:38:06	3.9	4.8	6.7
020	9/6/2019 13:53:06	3.8	5.0	6.0
021	9/6/2019 14:08:06	3.8	4.8	5.5
Peak		4.9	5.4	13.3
Min		0.0	0.7	1.2
Average		2.2	2.8	4.2